



JMS SOUTHEAST INCORPORATED CERTIFIES THIS DRAWING IS IN ACCORDANCE WITH JMS SOUTHEAST INCORPORATED ENGINEERING SPECIFICATIONS AND ACCURATELY REPRESENTS THE EQUIPMENT FURNISHED. THIS DRAWING IS THE SOLE PROPERTY OF JMS SE, AND MUST BE RETURNED UPON DEMAND IT MAY NOT BE REPRODUCED NOR THE INFORMATION CONTAINED THEREIN USED WITHOUT THE WRITTEN PERMISSION OF JMS SOUTHEAST INCORPORATED

Company Name:  
JMS Southeast Inc.

Reference Data:

Drawing #:

QD9067

Rev 0

Sensor	
Sensor	Type B Beaded Thermocouple, Single Element, Standard Limits [ASTM E230-11], Welded bead
Construction	24 AWG; 24" L
Attaching Device	None
Cold-End	Bare Ends; 2" Fiberglass Sleeve (Std) lead wire
JMS Part #	1HB1B24L24ZA

Project Name:

Generated By

Date

Initials

01/12/26

TOLERANCE TABLE  
PROBE (L) 0-48" ±18"  
48"-144" ±14"  
144" ±1" / ±2"  
WIRE (L) >24" -12" / ±1"  
>24" -12" / ±2"  
PROT. TUBE (P8U) ±14"  
FRACTIONAL DIM. ±1/8"  
DECIMAL DIM. ±1/8"  
XX ± 0.10"  
XXX ± 0.010"  
XXXX ± 0.005"  
UNLESS STATED OTHERWISE



**JMS SOUTHEAST, INC.**  
Temperature Measurement  
105 Temperature Lane  
Statesville, North Carolina 28677  
Phone: (704) 873-1835  
Fax: (704) 878-6166  
www.jms-se.com  
Email: sensors@jms-se.com

Not to Scale

Rev. 10-01-2015